

**DPL**  
Handling you with care



**DIPPED PRODUCTS  
PLC**

400, DEANS ROAD  
COLOMBO 10, SRILANKA  
PHONE: (011) 2 693964, 4 712748  
FAX: (011) 2 699018  
E-MAIL: postmast@dplgroup.com

Report Date: 2019/11/28

## CERTIFICATE OF CONFORMITY

**CUSTOMER** : VULKAN MEDICAL A.S.  
**PO NO** : 192000165  
**PRODUCT CODE** : 621U – 35.29 / CIH / P2  
**CUSTOMER STYLE** : JANA  
**LOT NUMBERS** : 20191026L2C , 20191027L2C , 20191027L2D , 20191028L2C ,  
20191025L2C

We hereby certify that the above product has been manufactured in accordance with a Quality Assurance Scheme certified to ISO 9001: 2015 International Standard.

The product conforms to European Norms EN374, EN 388 & EN 420 and has been inspected in accordance with ISO 2859-1 Sampling Plans Indexed by Acceptable Quality Level (AQL) for Lot-by-Lot Inspection and found to be conforming in respect of the following.

Parameter	Defectives	Inspection Level	AQL
Glove Quality	Major Defectives	General Inspection Level I	1.5
	Minor Defectives	General Inspection Level I	4.0
Packing Quality	Inner bags – Incorrect packing material, Sealing imperfections, Turned packing, Batch code imperfections, Incorrect packing method, Sizes mixed together, Single hands paired together.	S-3	4.0
	Outer bags - Incorrect packing material, Sealing imperfections, Turned packing, Incorrect quantity.	S3	4.0
	Outer cartons - Incorrect packing material, Batch code imperfections, Improper taping, Incorrect carton label, Incorrect quantity	S3	4.0
Dimensions	Thickness, Length	S2	4.0

  
-----  
Manager – Quality

2019/11/28  
-----  
Date